

Center for MR Research – 3T

Policy 4: Scanner Usage Following Non-Human Subjects

Following studies on non-human primates, the scanner must be clear for one hour.

PROCEDURE

- A. All surfaces must be cleaned and bleached including equipment, tables, scanner bed, etc.
- B. Human subjects are not to be scheduled for one hour after a primate has been scanned.
- C. The air circulation must be turned over four times after a primate has been scanned before a human subject may go into the scanner. It takes one hour for the air circulation to be turned over four times.

Sponsor: Director
Date: 05/20_rev3